

Animal and vegetable fats and oils — Determination of lead by direct graphite furnace atomic absorption spectroscopy

The European Standard EN 12193:2004 has the status of a
British Standard

ICS 67.200.10

National foreword

This British Standard is the official English language version of EN ISO 12193:2004. It is identical with ISO 12193:2004. It supersedes BS EN ISO 12193:2000 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee AW/307, Oil seeds animal and vegetable fats and oils, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this committee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Catalogue under the section entitled "International Standards Correspondence Index", or by using the "Search" facility of the BSI Electronic Catalogue or of British Standards Online.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, the EN ISO title page, the EN ISO foreword page, the ISO title page, pages ii and iii, a blank page, pages 1 to 7 and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

Amendments issued since publication

Amd. No.	Date	Comments
15083	17 March 2004	Corrected EN ISO foreword page
Corrigendum No. 1		

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 25 February 2004

© BSI 17 March 2004

ISBN 0 580 43492 3

English version

Animal and vegetable fats and oils - Determination of lead by
direct graphite furnace atomic absorption spectroscopy (ISO
12193:2004)

Corps gras d'origines animale et végétale - Détermination
de la teneur en plomb par spectrométrie d'absorption
atomique directe avec four en graphite (ISO 12193:2004)

Tierische und pflanzliche Fette und Öle - Bestimmung von
Blei durch direkte Graphitofen-
Atomabsorptionsspektrometrie (ISO 12193:2004)

This European Standard was approved by CEN on 19 February 2004.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

CORRECTED 2004-02-25

Foreword

This document (EN ISO 12193:2004) has been prepared by Technical Committee ISO/TC 34 "Agricultural food products" in collaboration with Technical Committee CEN/TC 307 "Oilseeds, vegetable and animal fats and oils and their by-products - Methods of sampling and analysis", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2004, and conflicting national standards shall be withdrawn at the latest by August 2004.

This document supersedes EN ISO 12193:1999.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 12193:2004 has been approved by CEN as EN ISO 12193:2004 without any modifications.

NINRETOITALAN
ADNATSDR
INTERNATIONAL
STANDARD

OSI
13912
ISO
12193

ceSond deition
-400210-20

Second edition
~~2004-02-01~~

**Animal and vegetable fats and oils —
Determination of lead by ~~direct~~ graphite
furnace ~~atomic absorption~~ spectroscopy**

*Corps gras d'origines animale et végétale — Détermination de la teneur
en plomb par spectrométrie d'absorption atomique directe avec four en
graphite*



Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 12193 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*.

This second edition cancels and replaces the first edition (ISO 12193:1994), which has been technically revised.

Animal and vegetable fats and oils — Determination of lead by direct graphite furnace atomic absorption spectroscopy

1 Scope

This International Standard specifies a method for the determination of trace amounts ($> 0,001$ mg/kg) of lead in all types of crude or refined edible oils and fats.

2 Principle

The oil or fat is atomized in a graphite furnace, with or without a platform, connected to an atomic absorption spectrometer which has been previously calibrated using standard solutions of an organo compound of lead. The metal content is measured from the observed absorbance at a wavelength of 283,3 nm.

3 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified, and distilled or demineralized water or water of equivalent purity.

3.1 Cyclohexane, analytical grade.

3.2 Lecithin, containing 2 % phosphorus.

3.3 Matrix modifier: 2 % (by mass/volume) lecithin solution prepared by dissolving 2 g of lecithin (3.2) in 100 ml of cyclohexane (3.1).

3.4 Blank oil, refined.

Any edible oil is suitable, stored in a metal-free polyethylene bottle. The lead content of the oil shall not be greater than 0,001 mg/kg.

3.5 Lead standard stock solution, $c(\text{Pb}) = 10$ mg/kg.

This is prepared by appropriate dilution of an organometallic standard with the blank oil (3.4).

NOTE A suitable standard is available from the following companies:¹⁾

Continental Oil Company, Ponca City, Oklahoma, USA (Conostan, 5 000 mg/kg);

Merck KGaA, Darmstadt, Germany [Certipur 115051, lead standard dissolved in oil, $c(\text{Pb}) = 1$ g/kg] in standard oil;

VWR International [Certipur 1.15051, lead standard dissolved in oil, $c(\text{Pb}) = 1$ g/kg] in standard oil;

VWR International [Specpure ALFA019527.AD, lead, Organic AAS standard solution, Specpure, $c(\text{Pb}) = 1$ 000 mg/g].

1) These are examples of suitable products available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.

3.6 Lead standard working solutions: $c(\text{Pb}) = 0,02 \text{ mg/kg}$, $0,05 \text{ mg/kg}$ and $0,100 \text{ mg/kg}$, prepared daily by diluting the 10 mg/kg stock solution (3.5) with blank oil (3.4).

3.7 Argon, of minimum purity 99,99 %.

4 Apparatus

4.1 Polyethylene or polypropylene capped bottles, 20 ml, metal free.

The polyethylene or polypropylene capped bottles are made metal free in the following way. Clean the bottles thoroughly with warm nitric acid, rinse with distilled water and dry the bottles in a drying oven at about $80 \text{ }^\circ\text{C}$.

4.2 Micropipettor, $20 \mu\text{l}$.

4.3 Pipettor tips.

4.4 Electric oven, regulated at $60 \text{ }^\circ\text{C} \pm 2 \text{ }^\circ\text{C}$.

4.5 Atomic absorption spectrometer, equipped with either "peak height" mode and printer, or "area" mode and pen-recorder (full-scale response in $0,2 \text{ s}$), together with the appropriate electrode-less discharge lamp (or hollow cathode lamp) and deuterium or Zeeman background corrector.

4.6 Graphite furnace atomizer, placed in the atomic absorption spectrometer (4.5), equipped with a control unit for temperature programming.

4.7 Graphite tube, normal (uncoated) or coated (pyrolytic).

4.8 Pyrolytic platform, in combination with uncoated or pyrolytically coated graphite tube.

Both atomization with and without the platform may be used. The accuracy and the sensitivity with the platform is twice as good as without the platform.

5 Sampling

Laboratory samples of at least $1,5 \text{ g}$ shall be stored in metal-free bottles (4.1).

It is important that the laboratory receive a sample that is truly representative and has not been damaged or changed during transport or storage.

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 5555.

6 Procedure

6.1 Treatment of samples, blanks and standards

6.1.1 Place all samples and standard working solutions (3.6) in the oven (4.4) set at $60 \text{ }^\circ\text{C}$.

6.1.2 Shake samples vigorously.

6.1.3 Weigh $5,0 \text{ g}$ of sample into a 20 ml bottle (4.1), together with $5,0 \text{ g}$ of matrix modifier (3.3), and mix thoroughly.

6.1.4 Carry out steps 6.1.2 and 6.1.3 also for the three standard working solutions (3.6) and the blank oil (3.4).

6.2 Preparation of apparatus

6.2.1 Switch on the atomic absorption spectrometer (4.5) and the background correction (D₂ or Zeeman).

6.2.2 In accordance with the manufacturer's instructions supplied with the spectrometer, adjust the lamp current, slit, wavelength and amplification. The required wavelength is 283,3 nm.

6.2.3 Optimize the position of the graphite furnace atomizer (4.6) in the atomic absorption spectrometer and set the required programme on the control unit of the furnace. Use the measuring parameters given in Tables 1 and 2. If available, place the platform (4.8) in the graphite tube.

6.2.4 Before each injection, pretreat the pipettor tip (4.3) by pipetting and then discarding 20 µl of cyclohexane.

Programmes for the graphite furnace atomizer are given in Tables 1 and 2.

Table 1 — Uncoated tube without platform, with maximum power heating and gas-stop

Step No.	Temperature °C	Ramp-time s	Hold-time s	Internal gas flow ml/min
1	100	10	20	300
2	650	60	40	300
3	1 900	0	5	0
4	2 700	1	3	50

Table 2 — Pyro-coated tube with platform, with maximum power heating and gas-stop

Step No.	Temperature °C	Ramp-time s	Hold-time s	Internal gas flow ml/min
1	200	10	20	300
2	650	60	40	300
3	1 700	0	5	0
4	2 700	1	3	50

6.3 Determination

6.3.1 Measurement of blank solution

Inject 20 µl of the blank solution, prepared according to 6.1.4, into the graphite furnace. Initiate the temperature programme and record the absorption.

6.3.2 Measurement of standard solutions

Inject 20 µl of the three standard solutions, prepared according to 6.1.4, into the graphite furnace and record the absorptions.

6.3.3 Measurement of sample solutions

Inject 20 µl of the sample solution, prepared according to 6.1.3, into the graphite furnace. Initiate the temperature programme and record the absorption.

7 Calculation and expression of results

7.1 Calculation

7.1.1 Measure the peak height on the recorder-chart or take the reading of the display or printer.

7.1.2 Draw a calibration curve by plotting the absorption of the three standard solutions (6.3.2), corrected for the blank (6.3.1), against their respective metal content.

NOTE With the use of sophisticated equipment, autocalibration can be applied.

7.1.3 Read the metal content of the sample from the calibration curve.

7.2 Expression of results

Express the results as milligrams per kilogram (to two significant figures).

8 Precision

8.1 Interlaboratory test

Details of an interlaboratory test on the precision of the method are summarized in Annex A. The values derived from this interlaboratory test may not be applicable to concentration ranges and matrices other than those given.

The method has been tested with liquid oil (soyabean oil) and with solid fat (cocoa butter).

8.2 Repeatability limit

The absolute difference between two independent single test results, obtained with the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will not in more than 5 % of cases be greater than the repeatability limit (r) as calculated from the formulae in Table 3.

Table 3 — Repeatability (r) and reproducibility limits (R)

	r	R
Pb in edible oil	0,19 M	0,30 M
Pb in cocoa butter	0,15 M	0,68 M

NOTE M is the corresponding mean concentration value.

8.3 Reproducibility limit

The absolute difference between two single test results, obtained with the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than the reproducibility limit (R) as calculated from the formulae in Table 3.

8.4 Trueness (bias)

The bias of the method was demonstrated in the collaborative study of the method (see Table A.1) to be negligible when used for the determination of concentration levels of lead in the range 0,01 mg/kg to 0,10 mg/kg.

8.5 Sensitivity

The sensitivity of the method is demonstrated by the low values for ρ and P at the low concentration levels studied (see Table A.1). The limit of detection is 0,001 mg/kg and the limit of determination is 0,01 mg/kg.

Interference by other elements is not to be expected, provided that the measurements are carried out at the wavelength specific for lead (283,3 nm).

9 Test report

The test report shall specify:

- a) all information necessary for the complete identification of the sample;
- b) the sampling method used, if known;
- c) the test method used, with reference to this International Standard;
- d) all operating details not specified in this International Standard, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- e) the test result(s) obtained, or, if the repeatability has been checked, the final quoted result obtained.

Annex A (informative)

Results of interlaboratory test

An interlaboratory test was carried out at the international level in 1988 by the IUPAC Commission on Oils, Fats and Derivatives, in which 20 laboratories participated, each obtaining two test results for each sample. The statistical results obtained (evaluated in accordance with ISO 5725:1986)²⁾ are summarized in Table A.1.

Table A.1 — Statistical results

	Sample					
	Soyabean oil			Cocoa butter		
	A	B	C	A	B	C
Number of laboratories retained after eliminating outliers	14	14	14	16	16	16
Number of outliers (laboratories)	6	6	6	4	4	4
Number of accepted results	112	112	112	128	128	128
Mean value (mg/kg sample)	0,087	0,053	0,022	0,087	0,054	0,026
True or accepted value (mg/kg)	0,085	0,050	0,020	0,085	0,050	0,020
Repeatability standard deviation, s_r (mg/kg)	0,0059	0,0034	0,0022	0,0039	0,0031	0,0013
Repeatability relative standard deviation (%)	6,8	6,4	10,1	4,5	5,7	5,0
Repeatability limit (r) ($2,8 \times s_r$)	0,017	0,0095	0,0062	0,011	0,0087	0,004
Reproducibility standard deviation, s_R (mg/kg)	0,0089	0,0057	0,0034	0,0180	0,0117	0,0073
Reproducibility relative standard deviation (%)	10,2	10,9	15,4	20,6	21,7	27,7
Reproducibility limit (R) ($2,8 \times s_R$)	0,025	0,016	0,009	0,050	0,033	0,020

2) ISO 5275:1986, *Precision of test methods — Determination of repeatability and reproducibility for a standard test method by inter-laboratory tests* (now withdrawn).

Bibliography

- [1] ISO 5555:2001, *Animal and vegetable fats and oils — Sampling*
- [2] HENDRIKSE P.W., SLIKKERVEER F.J., ZAALBERG J. and HAUTFENNE A. *Pure Appl. Chem.*, **60**, 1988, pp. 893
- [3] HENDRIKSE P.W., SLIKKERVEER F.J., FOLKERSMA A. and DIEFFENBACHER A. *Pure Appl. Chem.*, **63** (8), 1991, pp. 1185-1192 (Report of the collaborative study)
- [4] SLIKKERVEER F.J., BRAAD A.A. and HENDRIKSE P.W. *At. Spectr.*, **1**, 1980, p.30

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001.

Fax: +44 (0)20 8996 7001. Email: orders@bsi-global.com. Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: info@bsi-global.com.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001.

Email: membership@bsi-global.com.

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager. Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553. Email: copyright@bsi-global.com.